

CERTIFICATE OF ANALYSIS

ME0304_25984712/1

Product Methanol, dried (max. 0,005% H2O), for analysis, ExpertQ® (Karl Fischer)

Batch 25984712

Quality release date 24/07/2025

ME0304

Expiry date 07/2028

Analysis	Batch value	Specifications
Assay (G.C.)	99,99 %	≥ 99,8
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	0,7912 g/mL	0,790 - 0,792
Appearance	clear	clear
Colour (Hazen)	1	≤ 10
Solubility in water	passes test	passes test
Acidity	0,00006 meq/g	≤ 0,0003
Alkalinity	0,00001 meq/g	≤ 0,0002
Chlorides (Cl)	< 0,00005 %	≤ 0,00005
Sulfates (SO4)	< 0,0001 %	≤ 0,0001
Aluminium (Al)	< 0,5 ppm	≤ 0,5
Arsenic (As)	< 0,02 ppm	≤ 0,02
Barium (Ba)	< 0,1 ppm	≤ 0,1
Beryllium (Be)	< 0,02 ppm	≤ 0,02
Bismuth (Bi)	< 0,02 ppm	≤ 0,02
Boron (B)	< 0,02 ppm	≤ 0,02
Cadmium (Cd)	< 0,05 ppm	≤ 0,05
Calcium (Ca)	< 0,5 ppm	≤ 0,5
Chromium (Cr)	< 0,02 ppm	≤ 0,02
Cobalt (Co)	< 0,02 ppm	≤ 0,02
Copper (Cu)	< 0,02 ppm	≤ 0,02
Gallium (Ga)	< 0,02 ppm	≤ 0,02
Gold (Au)	< 0,02 ppm	≤ 0,02
Indium (In)	< 0,02 ppm	≤ 0,02
Iron (Fe)	< 0,1 ppm	≤ 0,1
Lead (Pb)	< 0,02 ppm	≤ 0,02
Lithium (Li)	< 0,05 ppm	≤ 0,05
Magnesium (Mg)	< 0,1 ppm	≤ 0,1
Manganese (Mn)	< 0,01 ppm	≤ 0,01
Molybdenum (Mo)	< 0,02 ppm	≤ 0,02
Nickel (Ni)	< 0,02 ppm	≤ 0,02
Platinum (Pt) ppm	< 0,05 ppm	≤ 0,05
Silver (Ag)	< 0,02 ppm	≤ 0,02
Thallium (Tl)	< 0,02 ppm	≤ 0,02
Tin (Sn)	< 0,1 ppm	≤ 0,1
Titanium (Ti)	< 0,02 ppm	≤ 0,02
Vanadium (V)	< 0,02 ppm	≤ 0,02
Zinc (Zn)	< 0,1 ppm	≤ 0,1
Zirconium (Zr)	< 0,02 ppm	≤ 0,02
Acetone (G.C.)	< 0,001 %	≤ 0,001

Results marked with # are out of specifications.
 This certificate does not exempt the user from checking the results upon receipt of the goods.
 Any copy of our CoA may be obtained from our website at www.scharlab.com

 Marta Martínez Casals
 Laboratory Technician



**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**ME0304_25984712/1**

Product	Methanol, dried (max. 0,005% H ₂ O), for analysis, ExpertQ® (Karl Fischer)	Batch	25984712
		Quality release date	24/07/2025
ME0304		Expiry date	07/2028

Analysis	Batch value	Specifications
Ethanol (G.C.)	< 0,005 %	≤ 0,1
Aldehydes and ketones (as C ₂ H ₅ CHO)	< 0,001 %	≤ 0,001
Acetaldehyde (CH ₃ CHO)	< 0,001 %	≤ 0,001
Formaldehyde	< 0,0001 %	≤ 0,0001
Substances reducing KMnO ₄	passes test	passes test
Substances darkened by H ₂ SO ₄	passes test	passes test
Residue on evaporation	0,0001 %	≤ 0,0005
Water (K.F.)	0,0048 %	≤ 0,005

For laboratory use only

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martínez Casals
Laboratory Technician