



Scharlab S.L.

Tel. int.: +34-93-7151811

Email: scharlab@scharlab.com

CERTIFICATE OF ANALYSIS

Product: **Methanol**, for GC residue analysis

Batch 16438516

ME0318

Quality release date 25/06/2015

Expiry date 6/2018

Analysis	Batch value	Specifications
assay (G.C.)	99,94 %	min. 99,9 %
identity (IR-spectrum)	passes test	passes test
density (20°/4°)	0,7913	0,790 - 0,792
residue on evaporation	< 0,0001 %	max. 0,0001 %
water (K.F.)	0,02 %	max. 0,03 %
Suitable for organohalogenated		
pesticide and dioxins, furans and PCBs		
residue analysis		
ECD, from 1,2,4-trichlorobenzene to		
decachlorobiphenyl, no peaks are		
obtained greater than 3 pg/ml as lindane.		
No peaks are obtained in vicinity of		
2,4,5-trichlorobiphenyl.		

This certificate does not release the user from their control upon receipt of the goods
You can get a copy of any of our COA from our web site: www.scharlab.com

M. Canet
Laboratory Manager