

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

Product	Methanol, for GC residue analysis	Batch	20033616
ME0318		Quality release date	25/07/2019
		Expiry date	07/2022
Analysis	Batch value	Specifications	
Assay (G.C.)	99,99 %	≥ 99,9	
Identity (IR-Spectrum)	passes test	passes test	
Density (20°/4°)	0,7914 g/mL	0,790 - 0,792	
Residue on evaporation	0,0001 %	≤ 0,0001	
Water (K.F.)	0,0200 %	≤ 0,03	

Suitable for organohalogenated pesticide and dioxins, furans and PCBs residue analysis.

ECD, from 1,2,4-trichlorobenzene to decachlorobiphenyl, no peaks are obtained greater than 3 pg/ml as lindane.

No peaks are obtained in vicinity of 2,4,5-trichlorobiphenyl.