

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**ME0318_21205116/1**

Product Methanol, for GC residue analysis

Batch 21205116

ME0318

Quality release date 20/10/2020

Expiry date 10/2023

Analysis	Batch value	Specifications
Assay (G.C.)	99,98 %	≥ 99,9
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	0,7918 g/mL	0,790 - 0,792
Residue on evaporation	0,0001 %	≤ 0,0001
Water (K.F.)	0,0095 %	≤ 0,03

Suitable for organohalogenated pesticide and dioxins, furans and PCBs residue analysis.

ECD, from 1,2,4-trichlorobenzene to decachlorobiphenyl, no peaks are obtained greater than 3 pg/ml as lindane.

No peaks are obtained in vicinity of 2,4,5-trichlorobiphenyl.

For laboratory use only