

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**ME0326_25727318/1**

Product Methanol, LC-MS, suitable for nitrosamine analysis

Batch 25727318

ME0326

Quality release date 09/04/2025

Expiry date 04/2030

Analysis	Batch value	Specifications
Assay (G.C.)	99,98 %	≥ 99,9
Identity (IR-Spectrum)	passes test	passes test
Density (20°/4°)	0,7916 g/mL	0,790 - 0,792
Acidity	0,00006 meq/g	≤ 0,0002
Alkalinity	0,00001 meq/g	≤ 0,0002
Aluminium (Al)	< 10 ppb	≤ 10
Barium (Ba)	< 5 ppb	≤ 5
Cadmium (Cd)	< 5 ppb	≤ 5
Calcium (Ca)	< 10 ppb	≤ 10
Chromium (Cr)	< 5 ppb	≤ 5
Cobalt (Co)	< 5 ppb	≤ 5
Copper (Cu)	< 5 ppb	≤ 5
Iron (Fe)	< 10 ppb	≤ 10
Lead (Pb)	< 5 ppb	≤ 5
Magnesium (Mg)	< 10 ppb	≤ 10
Manganese (Mn)	< 5 ppb	≤ 5
Nickel (Ni)	< 5 ppb	≤ 5
Potassium (K)	< 10 ppb	≤ 10
Silver (Ag)	< 5 ppb	≤ 5
Sodium (Na)	< 50 ppb	≤ 50
Tin (Sn)	< 5 ppb	≤ 5
Zinc (Zn)	< 10 ppb	≤ 10
Residue on evaporation	0,0002 %	≤ 0,0005
Water (K.F.)	0,0090 %	≤ 0,02
Nitrosamine analysis	passes test	passes test
Suitability for use in LC-MS	passes test	passes test
UV Spectroscopy in a 1,0 cm cell:		
min. transmission at 205 nm	36,81 %	≥ 20
min. transmission at 220 nm	84,78 %	≥ 60
min. transmission at 240 nm	96,83 %	≥ 90
min. transmission at 260 nm	99,39 %	≥ 98
max. absorbance at 205 nm	0,434 AU	≤ 0,699
max. absorbance at 220 nm	0,072 AU	≤ 0,222
max. absorbance at 240 nm	0,014 AU	≤ 0,046
max. absorbance at 260 nm	0,003 AU	≤ 0,009
Gradient grade (235 nm):		
maximum peak absorbance:	0,0003 AU	≤ 0,001
Gradient grade (254 nm):		
maximum peak absorbance:	0,0002 AU	≤ 0,0005

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martínez Casals
Laboratory Technician



Scharlab S.L.

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

ME0326_25727318/1

Microfiltered through membranes of pore diameter 0,1 µm.

For laboratory use only.