

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**PL0051_22566401/1**Product Silver nitrate, concentrated solution to prepare 1 l of solution 0,1 mol/l
(0,1 N)

Batch 22566401

Quality release date 04/03/2022

Expiry date 03/2025

PL0051

Analysis	Batch value	Specifications	±U
Concentrated solution	1 mol/L	1 mol/L	0,1 %

Preparation

This volumetric solution is prepared using gravimetric and volumetric procedures.

Contains 16,987g Silver nitrate

Measurement

This concentrated solution was checked using Scharlau's potassium chloride volumetric standard.

Storage and use

Concentrate to prepare 1 litre of solution according to the instructions printed in the box.

If needed, different concentrations may be obtained by dilution to a volume other than 1 litre.

Included in the box is an extra label to be stuck in the user's flask.

This concentrated solution in the ampoule is stable for 5 years from the date of manufacturing.

Preserve the unopened ampoule in the original package.

For laboratory use only