

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**PO0168_25796001/1**

Product	Potassium bromide, IR spectroscopy grade	Batch	25796001
PO0168		Quality release date	07/05/2025
		Expiry date	05/2030
Analysis	Batch value	Specifications	
Absorbance (2g pressed disk, 0,5cm layer thickness):			
3250-3750 cm ⁻¹	< 0,15	≤ 0,15	
1620-1640 cm ⁻¹	< 0,15	≤ 0,15	
2750-3100 cm ⁻¹	< 0,004	≤ 0,004	
Other abs 420-4000 cm ⁻¹	< 0,015	≤ 0,015	
Suitability for IR spectroscopy	passes test	passes test	

For laboratory use only

Results marked with # are out of specifications.
This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com

Marta Martínez Casals
Laboratory Technician