

**Scharlab S.L.**

Crtra.Polinya a Sentmenat, Km 8,2

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**SO0005_22138601/1**

Product Sodium, standard solution 1000 mg/l Na for AA (sodium nitrate in nitric acid 0,5 mol/l)

Batch 22138601

Quality release date 05/10/2021

Expiry date 10/2024

SO0005

Analysis	Batch value	Specifications	±U
Concentration	996,8 mg/L	995,0 - 1005,0	1,2

Preparation

This atomic absorption standard has been prepared using gravimetric and volumetric procedures.

Composition per litre is 3,69 g Sodium nitrate in Nitric acid 0,5 mol/l.

Traceability

This standard is traceable to Standard Reference Material from NIST.

SRM 3152a lot No. 010728 Sodium Standard Solution.

Uncertainty

It characterises the dispersion of the values that could be attributed to the mesurand. The limits of the expanded uncertainty are given at a confidence level of 95% (k=2).

Measurement

The batch value is determined by ICP spectrophotometric method at 20°C.

Batch value certified at the time of measurement.

Storage and use

For AA spectrophotometer calibration

If the product is stored and unopened, this solution is stable for 3 years from the date of manufacturing.

Once the bottle is opened, keep tightly closed at room temperature. Avoid exposure to light.

We suggest rejecting the solution six months after opening

This solution should be checked periodically by the user.

For laboratory use only