



Scharlab S.L.

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

SO0734_26213301/1

Product	Sodium thiosulfate, solution 0,002 mol/l (0,002 N)	Batch	26213301
SO0734		Quality release date	23/10/2025
		Expiry date	10/2027
Analysis	Batch value	Specifications	±U
Factor	1,000	0,995 - 1,005	0,002

Preparation

This volumetric solution is prepared using gravimetric and volumetric procedures
Contains preservative.

1ml = 0,0004964g $\text{Na}_2\text{S}_2\text{O}_3 \cdot 5\text{H}_2\text{O}$

Traceability

This solution was analyzed using a certified reference material (potassium iodate). The certified reference material is ISO 17034 accredited, measured according to ISO/IEC 17025 and traceable to the International System of Units by means of a Standard Reference Material from NIST: SRM® 136 (Potassium Dichromate (Oxidimetric Standard)).

Uncertainty

It characterises the dispersion of the values that could be attributed to the mesurand. The limits of the uncertainty are given at a confidence level of 95% (k=2).

Measurement

The batch value is determined by means of potentiometric method (Red-ox titration) at 20°C.

The equipments involved in the analysis of this solution are validated periodically.

Batch value certified at the time of measurement.

Storage and use

For volumetric titration.

If the product is stored and unopened, this solution is stable for 2 years from the date of manufacturing.

Keep tightly closed at room temperature. Avoid exposure to light and moisture.

We suggest shaking the bottle before use in order to homogenize the content.

The titre should be checked periodically by the user.

For laboratory use only