

## Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2 08181 Sentmenat

**\( +34937456400** 

# CERTIFICATE OF ANALYSIS SO0737 22291201/1

Product Sodium thiosulfate, solution 0,05 mol/l (0,05 N)

Batch 22291201

Quality release date 26/11/2021

Expiry date 11/2024

 Analysis
 Batch value
 Specifications
 ±U

 Factor
 1,000
 0,999 - 1,001
 0,001

#### Preparation

SO0737

This volumetric solution is prepared using gravimetric and volumetric procedures Contains preservative.  $1ml = 0.01241g \ Na_2S_2O_3 \cdot 5H_2O$ 

#### Traceability

This solution was analyzed using a certified reference material (potassium iodate). The certified reference material is ISO 17034 accredited, measured according to ISO/IEC 17025 and traceable to the International System of Units by means of a Standard Reference Material from NIST: SRM® 136 (Potassium Dichromate (Oxidimetric Standard)).

### Uncertainty

It characterises the dispersion of the values that could be attributed to the mesurand. The limits of the uncertainty are given at a confidence level of 95% (k=2).

#### Measurement

The batch value is determined by means of potentiometric method (Red-ox titration) at 20°C.

The equipment involved in the analysis of this solution are validated periodically.

Batch value certified at the time of measurement.

#### Storage and use

For volumetric titration.

If the product is stored and unopened, this solution is stable for 3 years from the date of manufacturing.

Keep tightly closed at room temperature. Avoid exposure to light and moisture.

We suggest shaking the bottle before use in order to homogenize the content.

The titre should be checked periodically by the user.

For laboratory use only