

Scharlab S.L.

Crtra.Polinya a Sentmenat, Km 8,2 08181 Sentmenat

**** +34937456400

☑ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS SO1007 23901401/2

23901401 Buffer solution pH = 7,00 (20 °C) (Potassium dihydrogen Batch Product phosphate/di-Sodium hydrogen phosphate) Quality release date 09/06/2023 06/2026 Expiry date SO1007 ±U **Batch value Analysis Specifications** 6,99 6,99 - 7,01 pН 0,01

Preparation

Standard buffer solutions are prepared using gravimetric and volumetric procedures according to DIN19266. Composition per litre is 3,5 g Potassium dihydrogen phosphate and 12,5 g di-Sodium hydrogen phosphate. Contains preservative.

Temperature dependence of the pH value

When calibrating your pHmeter at different temperatures than 20°C, refer to the table below to introduce accurate pH values.

T (°C) 0 7.13 5 7,07 10 7,05 15 7.02 20 7,00 25 6,98 30 6.98 35 6,96 40 6,95 45 6,95 6,95 50

Traceability

This pH buffer solution is traceable to the International System of Units through Standard Reference Materials® from NIST:

SRM 185 Potassium hydrogen phthalate,

SRM 186 Phosphate Buffers,

SRM 187 Sodium tetraborate,

SRM 189 Potassium tetroxalate and

SRM 2193 Calcium carbonate.

Uncertainty

It characterises the dispersion of the values that could be attributed to the measurand. The limits of the expanded uncertainty are given at a confidence level of 95% (k=2).

Measuremen

The batch value is determined by measurement with a combination glass electrode against multiple-point calibration according to DIN 19268. Batch value certified at the time of measurement.

Storage and use

For pH-meter calibration.

If product is stored and unopened, this solution is stable for 3 years from the date of manufacturing.

Once the bottle is opened, store tightly closed at room temperature. Avoid exposure to light.

We suggest rejecting the solution six months after opening

Never introduce the electrode in the bottle for measurements.

Never pour the used solution back in the bottle.

For laboratory use only.



Ricard Balladares Abella Laboratory Technician