

**Scharlab S.L.**

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

☎ +34937456400

✉ helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS**SO3923_26075801/4**

Product	Buffer solution pH = 6.86 (25 °C), according to ChP, GB/T, Ph. Eur., USP	Batch	26075801
		Quality release date	03/09/2025
		Expiry date	09/2028
SO3923			
Analysis	Batch value	Specifications	±U
pH	6,87	6,85 - 6,87	0,01

Preparation

Standard buffer solutions are prepared using gravimetric and volumetric procedures according to DIN19268.

Composition per litre is 3.55 g of anhydrous disodium hydrogen phosphate and 3.40 g of potassium dihydrogen phosphate.

Temperature dependence of the pH value

T (°C)	pH
0.....	6.98
5.....	6.95
10.....	6.92
15.....	6.90
20.....	6.88
25.....	6.86
30.....	6.85
35.....	6.84
40.....	6.84
45.....	6.83
50.....	6.83
55.....	6.83
60.....	6.84

Traceability

This pH buffer solution is traceable to the International System of Units through Standard Reference Materials® from NIST:

SRM 185 Potassium hydrogen phthalate, pH = 4,01 (25°C)

SRM 186 Phosphate Buffers, pH = 6,86 (25°C)

SRM 187 Sodium tetraborate, pH = 9,20 (25°C)

SRM 189 Potassium tetraoxalate, pH = 1,68 (25°C)

and SRM 2193 Calcium carbonate, pH = 12,45 (25°C)

Uncertainty

It characterises the dispersion of the values that could be attributed to the measurand. The limits of the expanded uncertainty are given at a confidence level of 95% (k=2).

This certificate does not exempt the user from checking the results upon receipt of the goods.
Any copy of our CoA may be obtained from our website at www.scharlab.com
This certificate modifies and cancels the content of the certificate with name SO3923_26075801/3

Marta Martinez Casals
Laboratory Technician



Scharlab S.L.

Gato Pérez, 33. Pol. Ind. Mas d'en Cisa

08181 Sentmenat

+34937456400

helpdesk@scharlab.com

CERTIFICATE OF ANALYSIS

SO3923_26075801/4

Measurement

The batch value is determined by measurement with a combination glass electrode against multiple-point calibration according to DIN 19268.

Batch value certified at the time of measurement.

Intended use

This reference material is intended for the calibration of pH meters and electrodes used for pH measurement or as a control sample for measuring the pH value.

For laboratory use only.

Storage

If product is stored and unopened, this solution is stable for 3 years from the date of manufacturing.

Once the bottle is opened, store tightly closed at room temperature. Avoid exposure to light. We suggest rejecting the solution six months after opening.

Never introduce the electrode in the bottle for measurements.

Never pour the used solution back in the bottle.

A handwritten signature in black ink, appearing to be "Marta", written over a horizontal line.