



**Scharlab S.L.**

Tel. int.: +34-93-7151811

Email: [scharlab@scharlab.com](mailto:scharlab@scharlab.com)

## CERTIFICATE OF ANALYSIS

Product: **Tetrabutylammonium hydroxide**, solution 0,1  
**TE0115** mol/l, buffered with phosphates, HPLC grade

Batch 17769501  
Quality release date 27/01/2017  
Expiry date 1/2020

| Analysis                                                     | Batch value | Specifications |
|--------------------------------------------------------------|-------------|----------------|
| pH (20 °C)                                                   | 7,60        | 7,4 - 7,6      |
| absorbance of a 0,005 M solution<br>in a 1 cm cell at 254 nm | 0,001 AU    | max. 0,02 AU   |

Remarks: The glass volumetric material used for the standardisation of the solution, is according to the National Bureau of Standards NBSIR-74-661 (dec. 1974).

The bottle must be tightly closed after each use. Should the product not to be used for a few weeks after opening for the first time, please verify the factor of the solution.